

UNCLASSIFIED

AD NUMBER
AD882534
NEW LIMITATION CHANGE
TO Approved for public release, distribution unlimited
FROM Distribution authorized to U.S. Gov't. agencies only; Test and Evaluation; 1 Mar 1971. Other requests shall be referred to Naval Weapons Lab., Dahlgren, VA 22448.
AUTHORITY
USNSWC ltr, 4 Mar 1975

THIS PAGE IS UNCLASSIFIED


NWL TECHNICAL REPORT TR-2245
August 1968

AD 882534

AD No. _____
DDC FILE COPY

MUZZLE FLASH MEASUREMENTS ON LARGE GUNS

Edward M. Lerner



Distribution limited to U.S. Gov't. agencies only:
for information; 7 MAR 1971 Other requests
for document must be referred to _____

U.S. NAVAL WEAPONS LABORATORY
DAHLGREN, VIRGINIA



Reproduction of this document outside the
boundaries of the U. S. Government must have
prior approval of the U. S. Naval Weapons
Laboratory.

20

THIS REPORT HAS BEEN DELIMITED
AND CLEARED FOR PUBLIC RELEASE
UNDER DOD DIRECTIVE 5200.20 AND
NO RESTRICTIONS ARE IMPOSED UPON
ITS USE AND DISCLOSURE.

DISTRIBUTION STATEMENT A

APPROVED FOR PUBLIC RELEASE;
DISTRIBUTION UNLIMITED.


Best Available Copy

NWL Technical Report TR-2245
August 1968

MUZZLE FLASH MEASUREMENTS
ON LARGE GUNS

by

Edward M. Lerner
Engineering and Evaluation Department



Distribution limited to U.S. Gov't. agencies only;
Text and illustrations; 1 MAR 1971 Other requests
for this document must be referred to —

EACH TRANSMITTAL OF THIS DOCUMENT OUTSIDE THE
AGENCIES OF THE U. S. GOVERNMENT MUST HAVE
PRIOR APPROVAL OF THE U. S. NAVAL WEAPONS
LABORATORY.


Dahlgren, Va 22448

FOREWORD

In the testing of propellant at the U. S. Naval Weapons Laboratory, relative muzzle flash intensities on individual rounds of large guns were recorded by human observers. Because of the human eye response variance from individual to individual and for differing lighting conditions, it has been impossible to compare flash intensities with a reasonable degree of confidence. In order to accurately compare propellant flash measured under varying conditions, a calibrated photodiode detection system simulating human eye response was developed.

This work was performed in the Standardization and Technical Services Branch of the Instrument Development Division of the Engineering and Evaluation Department.

Released by:



D. W. STONER
Director, Engineering and
Evaluation Department

ABSTRACT

A photodiode, modified to give a human eye response, was calibrated against a standard of spectral irradiance. The calibration procedure is explained in detail. This photodiode can then be used to measure, in absolute units, the muzzle flash intensities of large guns. The results of previous tests, regardless of ambient lighting conditions, can be compared.

A computer program, graphs and schematics are included.

CONTENTS

	<u>Page</u>
FOREWORD	i
ABSTRACT	ii
INTRODUCTION	1
SYSTEM DESCRIPTION	1
CALIBRATION	1
REDUCTION OF DATA	2
CONCLUSION	3
RECOMMENDATIONS	3

APPENDICES:

- A. Computer Program Listing
- B. Figures and Tables
- C. Distribution

INTRODUCTION

The intent of this project was to develop a calibrated system to simulate the human eye in spectral response (Figure 1) in the measurement of flash. A photodiode (EG & G SGD-444) in conjunction with a wide band-pass optical filter (IL/WB-510) was chosen. This combination closely matches the human eye response and eliminates response to the infrared radiation from the flash. The system measures in absolute units just what the human eye sees. In addition, the photodiode-filter combination offers many advantages over the human eye. These include a known and unvarying response curve, good linearity, and calibration against a standard of spectral irradiance.

SYSTEM DESCRIPTION

The photodiode and filter are mounted in a detector housing which is located on the bridge at Main Range and aimed just ahead of the muzzle of the gun being fired. If the flash is expected to be unusually intense, a calibrated neutral density filter, or an optical attenuator, is placed in front of the filter. The output of the diode is fed to a galvanometer in a CEC 5-133 recorder. The output of the detector is amplified and matched to the galvanometer through a Honeywell amplifier. The system schematic is shown in Figure 2.

Prior to the test, calibration steps are placed on the oscillograph record by a flash calibration panel.

CALIBRATION

1. Procedure

Calibration of the gun flash detector is based on a standard of spectral irradiance. The standard lamp, which is traceable to the National Bureau of Standards, is factory calibrated by the Eppley Laboratory, Inc. and checked by NWL.

Figure 4 shows the optical filter/detector spectral response curve. This plot is obtained by first measuring the spectral variation of irradiance by use of a detector of known spectral responsivity at a known position, the plane of calibration (See Figure 3). With this information the spectral responsivity of the combination can be determined.

Multiplication of the lamp output curve by the normalized combination curve yields the spectral responsivity curve. Actual values are shown in Table 1. It is this curve which must be integrated to obtain the combination power responsivity (mv.watt^{-1}). A computer program has been written that permits the evaluation of this parameter. The program is included in Appendix A.

Calibrated neutral density filters (optical attenuators) are placed between the standard lamp and the detector to check linearity (Figure 5). Since the standard lamp obeys the inverse square law for distances greater than 43 cm, linearity can also be checked by changing the distance between lamp and detector. For instance, doubling the distance will cut the irradiance by 75%, while tripling will produce an 89% reduction. This method, however, is difficult to implement and is used only as a check of the neutral density filters.

2. Calibration Panel:

From the previously described procedure, the detector responsivity was determined. The calibration panel (Figure 6) places known steps of voltages on the oscillograph record (See Figure 7).

3. Calibration Uncertainty:

A survey of the available optical instruments now on the market will give ample evidence of the poor accuracy to be expected in radiometry today compared to the accuracies obtained in other fields of measurement; the calibration accuracy of the standard of spectral irradiance is $\pm 5\%$. Although this uncertainty drops to $\pm 2\%$ in the infrared region, it is highest in the visible spectrum.

The actual response curve obtained is shown in Figure 4. The error produced by the perturbation at 800 nm was calculated to be less than $\pm 3\%$. This is well within the uncertainty due to the standard of spectral irradiance.

The computer program was used to eliminate error due to integrating the calibration curve by mechanical methods. A sample problem used for a test showed a computer error of less than $\pm 0.6\%$.

REDUCTION OF DATA

From the calibration steps on the oscillograph record the values of flash irradiances can be determined from the trace deflection by the following relation.

$$E = D \cdot \Delta d \cdot R_p^{-1}$$

where E = irradiance in $\text{uw} \cdot \text{cm}^{-2}$

D = deflection responsivity in $\text{mv} \cdot \text{mm}^{-1}$

Δd = deflection above ambient in mm

R_p = detector responsivity $\text{mv} \cdot \text{cm}^2 \cdot \text{uw}^{-1}$

The deflection Δd is the peak deflection due to flash minus the

deflection due to ambient light as can be seen in Figure 8.

The value reported is the peak radiant flux density (E) incident on the detector face corrected to a 100 foot reference distance through use of the inverse square law.

CONCLUSION

Using the calibrated photodiode system, flash intensities can be measured in absolute units. This enables the intensities of individual rounds in a given test to be accurately compared. The results of previous tests, regardless of ambient lighting conditions, can also be compared.

RECOMMENDATIONS

Data acquisition and reduction can be done more quickly through use of peak reading meters rather than paper recorders. If the requirement for flash measurements becomes extensive, however, the system will have to be converted to a form suitable to automatic data processing without manual record reading.

APPENDIX A

COMPUTER PROGRAM LISTING

```

H      TYPE,COMPILEGO,FORTRAN,PM
T      SUBTYPE,FORTRAN,LMAP,PUNCH
1  FORMAT (10, 5X, F6.3)
2  FORMAT (10,2X,F10.3, 5X, F7.3)
3  FORMAT (8F10.3)
4  FORMAT(9H FOR THE ,10,26H SUBINTERVAL, THE AREA IS ,F14.4)
5  FORMAT(35H THE VALUE OF THE INTEGRAL FOR THE ,10,16H SET OF DATA 1
   IS ,F14.4)
   DIMENSION X(501), Y(501), XNCRMT(3001), YNCRMT(3001)
   INTEGER SUBIN
   REAL INTGRL
   READ 1, NOSETS, H
   DO 11 I = 1, NOSETS
     INTGRL = 0.0
     READ 2, ICARD, FIRSTX, D
     READ 3, (Y(J), J = 1, ICARD)
     DO 6 J = 1, ICARD
7  X(J) = FIRSTX + FLOAT(J - 1) * D
     NXNCRT = 2.0 * D / H + 1.0
     NSUBIN = ((ICARD - 3) / 2 + 1)
     DO 10 K = 1, NSUBIN
       XL11 = X(2 * K - 1)
       XL12 = X(2 * K)
       XL13 = X(2 * K + 1)
       YL11 = Y(2 * K - 1)
       YL12 = Y(2 * K)
       YL13 = Y(2 * K + 1)
       DO 7 L = 1, NXNCRT
         XNCRMT(L) = XL11 + FLOAT(L - 1) * H
7  YNCRMT(L) = ((XNCRMT(L) - XL12) * (XNCRMT(L) - XL13) * YL11) / ((
1  XL11 - XL12) * (XL11 - XL13)) + ((XNCRMT(L) - XL11) * (XNCRMT(L) -
2  XL13) * YL12) / ((XL12 - XL11) * (XL12 - XL13)) + ((XNCRMT(L) -
3  XL11) * (XNCRMT(L) - XL12) * YL13) / ((XL13 - XL11) * (XL13 - XL12
4  ))
       Y0 = YNCRMT(1)
       YN = YNCRMT(NXNCRT)
       YEVEN = 0.0
       YODD = 0.0
       LASTEX = NXNCRT - 2
       LASTOX = NXNCRT - 1
       DO 8 M1 = 3, LASTEX, 2
8  YEVEN = YEVEN + YNCRMT(M1)
       DO 9 M2 = 2, LASTOX, 2
9  YODD = YODD + YNCRMT(M2)
       AXL1L3 = H / 3.0 * (Y0 + 4.0 * YODD + 2.0 * YEVEN + YN)
       PRINT 4, K, AXL1L3
10  INTGRL = INTGRL + AXL1L3
       PRINT 5, I, INTGRL
     RETURN
   END
   989

```

APPENDIX B

FIGURES AND TABLES

TABLE 1
"Spectral Power Distribution"

nm	PD	nm	PD	nm	PD
400	0.016	535	7.59	670	2.29
405	0.030	540	8.27	675	1.82
410	0.048	545	9.10	680	1.44
415	0.067	550	9.79	685	1.36
420	0.085	555	10.6	690	1.05
425	0.112	560	11.2	695	1.07
430	0.148	565	11.8	700	0.873
435	0.231	570	12.4	705	0.993
440	0.257	575	12.8	710	1.00
445	0.318	580	13.0	715	1.01
450	0.420	585	13.0	720	0.793
455	0.534	590	12.9	725	0.919
460	0.692	595	12.7	730	0.696
465	0.802	600	12.0	735	0.817
470	1.08	605	11.6	740	0.823
475	1.25	610	10.8	745	0.709
480	1.51	615	9.94	750	0.714
485	1.80	620	8.91	755	0.599
490	2.18	625	8.13	760	0.602
495	2.52	630	7.34	765	0.729
500	2.98	635	6.35	770	0.486
505	3.47	640	5.52	775	0.488
510	4.04	645	4.87	780	0.739
515	4.63	650	4.02	785	0.738
520	5.21	655	3.63	790	0.742
525	5.97	660	2.86	795	0.622
530	6.64	665	2.61	800	0.622

Where nm = wavelength

PD = power density in (uw.cm^{-2})




FIGURE 1

NORMALIZED HUMAN EYE RESPONSE




FIGURE 2




FIGURE 3

ACTUAL RESPONSE CURVE
OF DETECTOR




FIGURE 4

CALIBRATION PROCEDURE




FIGURE 5

The diagram shows a circuit with a 22 1/2 VOLT battery connected to a series of resistors R_1 , R_2 , R_3 , and R_4 . A switch is connected to the junction between R_3 and R_4 . The other end of the switch is connected to a resistor R_L , which is then connected to terminal J2. Terminal J2 is also connected to a resistor R_D and then to terminal 2. Terminal 2 is labeled 'FROM DETECTOR'. Terminal 1 is connected to a common ground. A relay is connected between terminal 1 and the junction between R_1 and R_2 . The relay is labeled 'RELAY'. The circuit is also connected to terminal J1, which is labeled 'TO GALVANOMETER'. A ground symbol is shown near terminal J1.




FIGURE 6




FIGURE 7




FIGURE 8

UNCLASSIFIED

Security Classification

DOCUMENT CONTROL DATA - R & D

Security Classification of title, body, of abstract and indexing annotation must be entered when the overall report is classified

1. ORIGINATING ACTIVITY (Corporate author)		2a. REPORT SECURITY CLASSIFICATION	
Naval Weapons Laboratory Dahlgren, Virginia 22448		UNCLASSIFIED	
3. REPORT TITLE		2b. GROUP	
MUZZLE FLASH MEASUREMENTS ON LARGE GUNS			
4. DESCRIPTIVE NOTES (Type of report and inclusive dates)			
5. AUTHOR(S) (First name, middle initial, last name)			
Edward M. Lerner			
6. REPORT DATE	7a. TOTAL NO. OF PAGES	7b. NO. OF REFS	
August 1968			
8a. CONTRACT OR GRANT NO.	9a. ORIGINATOR'S REPORT NUMBER(S)		
b. PROJECT NO.	TR-2245		
c.	9b. OTHER REPORT NO(S) (Any other numbers that may be assigned this report)		
d.			
10. DISTRIBUTION STATEMENT			
Distribution limited to U.S. Gov't. agencies only; Test and Evaluation; March 1971. Other requests for this document must be referred to Commander, Naval Weapons Laboratory, Dahlgren, Virginia 22448.			
11. SUPPLEMENTARY NOTES		12. SPONSORING MILITARY ACTIVITY	
13. ABSTRACT			
<p>A photodiode, modified to give a human eye response, was calibrated against a standard of spectral irradiance. The calibration procedure is explained in detail. This photodiode can then be used to measure, in absolute units, the muzzle flash intensities of large guns. The results of previous tests, regardless of ambient lighting conditions, can be compared.</p> <p>A computer program, graphs and schematics are included.</p>			

DD FORM 1473 (PAGE 1)

5/76 0101-307-6891

UNCLASSIFIED

Security Classification